

Certification of Counting Chambers

cell depths of 0.100 m which are calibrated w certified annually by T	per manufactured at Hausser Scientifi m +/-2%. The Counting Chambers and ith a Mahr Milliness Comparator Indicates and with an accuracy of +/001mm on meets the mil standard #45662A.	re measured on a pneumatic gauge, ator Gage. Our indicators are
Calibrated by:	Francis X. Hughes, III	
Title:	Production Manager	
Signature:	Francis X. Hughes, III	
Thank you,		
HAUSSER SCIENTIF	IC CO.	