



Certification of Counting Chambers

Item # 3200 Description: Hausser Bright-Line Phase Hemacytometer

Each Counting Chamber manufactured at Hausser Scientific Company is measured for proper cell depths of 0.100 mm +/-2%. The Counting Chambers are measured on a pneumatic gauge, which are calibrated with a Mahr Milliness Comparator Indicator Gage. Our indicators are certified annually by Trescal with an accuracy of +/- .001mm, and repeatability of +/- 1/3 of graduation. Certification meets the mil standard #45662A.

Calibrated by: Francis X. Hughes, III

Title: Production Manager

Signature: Francis X. Hughes, III

Thank you,

HAUSSER SCIENTIFIC CO.